



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: EX320

Description: Multimeter + Voltage Detector (Autoranging)

Date of Issue: 08-Dec-17

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 2014/30/EU
Report Number: WT148002947
Report Date of Issue: 1/16/2015

Standards:

EN 61326-1:2013
EN 61326-2-2:2013

LVD Directive 2014/35/EU
Report Number: E201687-D1007-1/A0/C0-(M)
Report Date of Issue: 7/31/2017

Standards:

EN 61010-1:2010
IEC 61010-1 (ed.3)
IEC 61010-2-033 (ed.1)

RoHS Directive: 2011/65/EU

Standard:

EN 50581:2012

The test reports show that the product fulfills the requirement in the EC Low Voltage Directive, EMC Directive, and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC Low Voltage Directive, EMC Directive, and RoHS Directive.

Mark Sultzbach / QA Administrator
(for Tony Campagna / Quality & Engineering Services Director)