



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: MA220
Description: 400A AC/DC Clamp Meter
Date of Issue: 30-Mar-15

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 2004/108/EC
Report Number: WT138002105
Report Date of Issue: 8/27/2013

Standards:

EN 61326-1:2013
EN 61326-2-2:2013
IEC 61326-1:2012
IEC 61326-2-2:2012

LVD Directive 2006/95/EC
Report Number: WT133005540
Report Date of Issue: 4/29/2014

Standards:

EN 61010-1: 2010
EN 61010-2-032: 2012
EN 61010-2-033:2012

RoHS Directive: 2011/65/EU

Standard:

EN 50581:2012

The test reports show that the product fulfills the requirement in the EC Low Voltage Directive, EMC Directive, and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC Low Voltage Directive, EMC Directive, and RoHS Directive.



Tony Campagna
Director of Quality Services