



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: EX720

Description: 800A True RMS AC Clamp Meter with Type K Temperature

Date of Issue: 11-Jul-16

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 89/336/EEC
Report Number: WT58000067
Report Date of Issue: 1/31/2005

Standards:

EN 61326:1997+A1:1998:+A2:2001+A3:2003

LVD Directive 2014/35/EU
Report Number: 17015373 004
Report Date of Issue: 8/7/2013

Standards:

EN 61010-1:2010
EN 61010-2-032:2012
EN 61010-031:2002+A1
EN 61010-2-033:2012

RoHS Directive: 2011/65/EU

Standard:

EN 50581:2012

The test reports show that the product fulfills the requirement in the EC Low Voltage Directive, EMC Directive, and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC Low Voltage Directive, EMC Directive, and RoHS Directive.

A handwritten signature in black ink, appearing to read "Mark Sultzbach", written over a horizontal line.

Mark Sultzbach / QA Administrator
(for Tony Campagna / Director of Quality Services)