



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: EX850

Description: Clamp Meter, 1000A, CAT IV + IR - Droid

Date of Issue: 17-Oct-14

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 2004/108/EC
Report Number: WT148000585
Report Date of Issue: 4/15/2014

Standards:

EN 61326-1:2013
EN 61326-2-2:2013

LVD Directive: 2006/95/EC
Report Number: WT133003616
Report Date of Issue: 7/28/2014

Standards:

EN 61010-1:2010
EN 61010-2-032:2012
EN 61010-2-033:2012

R&TTE Directive: 1999/5/EC
Report Number: WT148000585
WT148000586
Report Date of Issue: 4/15/2014

Standards:


ETSI EN 301 489-1 V1.9.2 (2011-09)
ETSI EN300 328 V1.7.1 (2006-10)
EN 62479:2010
ETSI EN 301 489-17 V2.2.1 (2012-09)

RoHS Directive: 2011/65/EU

Standard:

EN 50581:2012

The test reports show that the product fulfills the requirement in the EC Low Voltage Directive, EMC Directive, R&TTE Directive, and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC Low Voltage Directive, EMC Directive, R&TTE Directive, and RoHS Directive.


Mark Sultzbach / QA Administrator
(for Tony Campagna / Director of Quality Services)