



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: IR300UV

Description: IR Thermometer with UV Leakage Detector

Date of Issue: 25-Aug-16

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 2014/30/EU
Report Number: TEEX1606147
Report Date of Issue: 6/24/2016

Standards:

EN 55011:2009/A1:2010 Group 1 (Class B)	EN 61000-4-2:2009
EN 61000-3-2:2006+A1:2009+A2:2009	EN 61000-4-3:2006+A1:2008+A2:2010
EN 61000-3-3:2013	EN 61000-4-4:2012
EN 61326-1:2013	EN 61000-4-5:2006
EN 61326-2-1:2013	EN 61000-4-6:2014
EN 61326-2-2:2013	EN 61000-4-8:2010
	EN 61000-4-11:2004

RoHS Directive: 2011/65/EU

Standard:

EN 50581:2012

The test reports show that the product fulfills the requirement in the EC EMC Directive and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC EMC Directive and RoHS Directive.

Mark Sultzbach
(for Tony Campagna / Director of Quality Services)